

LEMBAR  
HASIL PENILAIAN SEJAWAT SEBIDANG ATAU *PEER REVIEW*  
KARYA ILMIAH : JURNAL ILMIAH

Judul Artikel Ilmiah : Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)  
 Nama Penulis : **Harto Tanujaya**, Hirofumi Shintaku, Dai Kitagawa, Adianto, Susilodinata & Satoyuki Kawano  
 Jumlah Penulis : 6 (enam)  
 Status Pengusul : Penulis Pertama  
 Identitas Buku Ilmiah : a. Nama Jurnal : Journal of Engineering and Technological Sciences  
 b. Nomor ISSN : ISBN 2337-5779, E-ISSN : 2338-5502  
 c. Vol. No. Bln. Th. : No.1, Volume 45, Januari 2013  
 d. Penerbit : ITB Journal  
 e. Jumlah halaman : 12 halaman  
 f. Alamat Web Jurnal : <http://journals.itb.ac.id/index.php/jets/article/view/622/0>  
 g. DOI Artikel : <http://dx.doi.org/10.5614%2Fj.eng.technol.sci.2013.45.1.5>


Kategori Publikasi Buku Ilmiah (beri (√) pada kategori yang tepat)

Jurnal Ilmiah Internasional Bereputasi  
 Jurnal Ilmiah Internasional  
 Jurnal Ilmiah Nasional Terakreditasi  
 Jurnal Ilmiah Nasional Tidak Terakreditasi  
 Jurnal Ilmiah Terindex di DOAJ/CABI/COPERNICUS/Lainnya

Hasil Penilaian *Peer Review*

Komponen Yang Dinilai	Nilai Maksimal Jurnal Ilmiah (isi di kolom yang sesuai)					Nilai Akhir <i>peer</i> Yang Diperoleh
	Internasional Bereputasi	Internasional	Nasional Terakreditasi	Nasional Tidak Terakreditasi	Nasional Terindex DOAJ dll	
Kelengkapan dan kesesuaian unsur isi <i>prosiding</i> (10%)	90% x 10% x 40					3,6
Ruang lingkup dan kedalaman pembahasan (30%)	93% x 30% x 40					11,16
Kecukupan & kemutakhiran data/informasi dan metodologi (30%)	92% x 30% x 40					11,04
Kelengkapan unsur & kualitas penerbit (30%)	95% x 30% x 40					11,4
Nilai <i>peer</i> Maksimal (100%)	40					37,2
Kontribusi Pengusul; (nilai akhir <i>peer</i> x bobot penulis pertama) = <b>37,2 x 60% = 22,32</b>						<b>22,32</b>
Komentar/Usulan <i>Peer Review</i> : (Terlampir hal. 2)	1. Tentang kelengkapan dan kesesuaian unsur; 2. Tentang ruang lingkup dan kedalaman pembahasan; 3. Kecukupan dan kemutakhiran data/informasi dan metodologi; 4. Kelengkapan unsur dan kualitas penerbit; 5. Indikasi Plagiasi; 6. Kesesuaian Bidang Ilmu: <i>Terlampir</i>					

Jakarta, 18.12.2019  
 Penilai I



(Prof. Dr. Ir. Agustinus Purna Irawan)  
 NIDN/NIP : 0328087102 / 10398021  
 Jabatan/Pangkat/Bidang Ilmu: Professor/IVC/Teknik Mesin  
 Unit Kerja: Fakultas Teknik – Universitas Tarumanagara

KOMENTAR  
PEER REVIEW

1. Tentang kelengkapan dan kesesuaian unsur:

Artikel dengan judul **Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)**, ditulis secara benar sesuai dengan standar penulisan artikel ilmiah yang memuat pendahuluan, metode/peralatan yang digunakan, pengambilan data dan data, analisa dan kesimpulan.

2. Tentang ruang lingkup dan kedalaman pembahasan:

Artikel dengan judul **Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)**, membahas studi kasus secara eksperimental mengenai getaran sebuah membran untuk trapezoidal ABMP dengan spesifik dan mudah dipahami.

3. Kecukupan dan kemutakhiran data/informasi dan metodologi:

Metodologi terstruktur dan jelas, data dan referensi yang diambil up to date.

4. Kelengkapan unsur dan kualitas penerbit:

Editorial board dan reviewer untuk Jurnal "**Journal of Engineering and Technological Sciences**" jelas dan terstruktur dan dapat diakses online. Diterbitkan oleh **ITB** dan dapat dilacak melalui daring, bereputasi, terindeks oleh **Scopus dan berimpact factor, Q2, dengan SJR (2018) 0,21, dan H index 10**. Jurnal ber ISSN/ISBN.

5. Indikasi Plagiasi:

Artikel dengan judul **Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)** yang diterbitkan oleh ITB Journal Bandung, dan dapat dibaca secara daring melalui <http://journals.itb.ac.id/index.php/jets/article/view/622/0>, dan DOI <http://dx.doi.org/10.5614%2Fj.eng.technol.sci.2013.45.1.5> tidak ditemukan indikasi plagiasi dengan tingkat kesamaan menggunakan software **Turnitin sebesar 6 %**.

6. Kesesuaian Bidang Ilmu:

Artikel tersebut membahas tentang studi pendekatan mekanika pada getaran sebuah membran dan ada Linieritas keilmuan dengan pengusul.

Jakarta, 18.12.2019  
Penilai I



(Prof. Dr. Ir. Agustinus Purna Irawan)  
NIDN/NIP : 0328087102 / 10398021  
Jabatan/Pangkat/Bidang Ilmu: Professor/IVC/Teknik Mesin  
Unit Kerja: Fakultas Teknik – Universitas Tarumanagara



LEMBAR  
HASIL PENILAIAN SEJAWAT SEBIDANG ATAU PEER REVIEW  
KARYA ILMIAH : JURNAL ILMIAH

Judul Artikel Ilmiah : Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)  
 Nama Penulis : **Harto Tanujaya**, Hirofumi Shintaku, Dai Kitagawa, Adianto, Susilodinata & Satoyuki Kawano  
 Jumlah Penulis : 6 (enam)  
 Status Pengusul : Penulis Pertama  
 Identitas Buku Ilmiah : a. Nama Jurnal : Journal of Engineering and Technological Sciences  
 b. Nomor ISSN : ISBN 2337-5779, E-ISSN : 2338-5502  
 c. Vol. No. Bln. Th. : No.1, Volume 45, Januari 2013  
 d. Penerbit : ITB Journal  
 e. Jumlah halaman : 12 halaman  
 f. Alamat Web Jurnal : <http://journals.itb.ac.id/index.php/jets/article/view/622/0>  
 g. DOI Artikel : <http://dx.doi.org/10.5614%2Fj.eng.technol.sci.2013.45.1.5>

Kategori Publikasi Buku Ilmiah (beri (√) pada kategori yang tepat)

Jurnal Ilmiah Internasional Bereputasi  
 Jurnal Ilmiah Internasional  
 Jurnal Ilmiah Nasional Terakreditasi  
 Jurnal Ilmiah Nasional Tidak Terakreditasi  
 Jurnal Ilmiah Terindex di DOAJ/CABI/COPERNICUS/Lainnya

Hasil Penilaian Peer Review

Komponen Yang Dinilai	Nilai Maksimal Jurnal Ilmiah (isi di kolom yang sesuai)					Nilai Akhir Yang Diperoleh
	Internasional Bereputasi	Internasional	Nasional Terakreditasi	Nasional Tidak Terakreditasi	Nasional Terindex DOAJ dll	
Kelengkapan dan kesesuaian unsur isi <i>prosiding</i> (10%)	4					3,76
Ruang lingkup dan kedalaman pembahasan (30%)	12					11,4
Kecukupan & kemutakhiran data/informasi dan metodologi (30%)	12					11,04
Kelengkapan unsur & kualitas penerbit (30%)	12					11,64
Total = 100%	40					37,84
Kontribusi Pengusul; (nilai akhir <i>peer</i> x bobot penulis pertama) = 37,84 x 60% = 22,704						22,704
Komentar/Usulan Peer Review:	1. Tentang kelengkapan dan kesesuaian unsur: 2. Tentang ruang lingkup dan kedalaman pembahasan; 3. Kecukupan dan kemutakhiran data/informasi dan metodologi; 4. Kelengkapan unsur dan kualitas penerbit: 5. Indikasi Plagiasi: 6. Kesesuaian Bidang Ilmu: <i>Terlampir</i>					

Jakarta, 11-11-2019  
 Penilai

  
 (Dr. Ir. M. Sobron Yamin L., M.Sc.)  
 NIDN/NIP : 0114056705 / 10311009  
 Jabatan/Pangkat/Bidang Ilmu: Lektor Kepala/IV/Teknik Mesin  
 Unit Kerja: Fakultas Teknik – Universitas Tarumanagara

<p>KOMENTAR PEER REVIEW</p>	<p>1. Tentang kelengkapan dan kesesuaian unsur:</p> <p>Artikel <b>Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)</b>, ditulis sesuai dengan kaidah penulisan artikel ilmiah yang meliputi pendahuluan, metode/alat, data dan analisa serta kesimpulan.</p> <p>2. Tentang ruang lingkup dan kedalaman pembahasan:</p> <p>Artikel <b>Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)</b>, membahas studi kasus secara eksperimental mengenai getaran sebuah membran untuk trapezoidal ABMP dengan kedalaman pembahasan yang spesifik.</p> <p>3. Kecukupan dan kemutakhiran data/informasi dan metodologi:</p> <p>Data yang diambil dan digunakan untuk analisa dan referensi tergolong baru dan mutakhir, dengan susunan metodologi yang baik.</p> <p>4. Kelengkapan unsur dan kualitas penerbit:</p> <p>Editor dan reviewer Jurnal "<b>Journal of Engineering and Technological Sciences</b>" tersusun dan terorganisir, ber ISSN/ISBN dan dapat diakses online. Diterbitkan oleh ITB dan dapat dilacak melalui daring, bereputasi, terindeks oleh <b>Scopus dan berimpact factor, Q2, dengan SJR (2018) 0,21, dan H index 10.</b></p> <p>5. Indikasi Plagiasi:</p> <p>Artikel <b>Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)</b> diterbitkan oleh ITB Journal Bandung, dan dapat dibaca secara daring melalui <a href="http://journals.itb.ac.id/index.php/jets/article/view/622/0">http://journals.itb.ac.id/index.php/jets/article/view/622/0</a>, dan DOI <a href="http://dx.doi.org/10.5614%2Fj.eng.technol.sci.2013.45.1.5">http://dx.doi.org/10.5614%2Fj.eng.technol.sci.2013.45.1.5</a> tidak ditemukan indikasi plagiasi.</p> <p>6. Kesesuaian Bidang Ilmu:</p> <p>Artikel tersebut membahas tentang studi pendekatan mekanika pada getaran sebuah membran dan ada kesesuaian dan linieritas keilmuan dengan pengusul.</p>
---------------------------------	--

Jakarta, 14-11-2019  
Penilai I



(Dr. Ir. M. Sobron Yamin L., M.Sc.)

NIDN/NIP : 0114056705 / 10311009

Jabatan/Pangkat/Bidang Ilmu: Lektor Kepala/IV/Teknik Mesin

Unit Kerja: Fakultas Teknik – Universitas Tarumanagara

# Journal of Engineering and Technological Sciences

**Country** [Indonesia](#) - [SJR Ranking of Indonesia](#)

**Subject Area and Category** [Engineering](#)  
[Engineering \(miscellaneous\)](#)

**Publisher** [Institut Teknologi Bandung \(ITB\)](#)

**Publication type** [Journals](#)

**ISSN** [23375770](#)

**Coverage** [2013-ongoing](#)


# 10

H Index

**Scope** Journal of Engineering and Technological Sciences welcomes full research articles in the area of Engineering Sciences from the following subject areas: Aerospace Engineering, Biotechnology, Chemical Engineering, Civil Engineering, Electrical Engineering, Engineering Physics, Environmental Engineering, Industrial Engineering, Information Engineering, Mechanical Engineering, Material Science and Engineering, Manufacturing Processes, Microelectronics, Mining Engineering, Petroleum Engineering, and other application of physical, biological, chemical and mathematical sciences in engineering. Authors are invited to submit articles that have not been published previously and are not under consideration elsewhere.

- [Homepage](#)
- [How to publish in this journal](#)
- [Contact](#)
- [Join the conversation about this journal](#)

## Quartiles



## SJR



## Total Cites



## Self-Cites



## External Cites per Doc



## Cites per Doc



## Citations per document



## % International Collaboration



## Citable documents



## Cited documents



## Uncited documents



Journal of Engineering and Technological Sciences



Show this widget in your own website

Just copy the code below and paste within your html code:

`<a href="https://www.scimago.com" data-bbox="491 857 593 868">`





Journal Content

Search  
All  
Search

Browse

- By Issue
- By Author
- By Title
- Other Journals

Information

- For Readers
- For Authors
- For Librarians

Current Issue

4100	1.0
437	1.0
452	1.0

Home > Vol 51, No 5 (2019)

Journal of Engineering and Technological Sciences



Journal of Engineering and Technological Sciences welcomes full research articles in the area of Engineering Sciences from the following subject areas: Aerospace Engineering, Biotechnology, Chemical Engineering, Civil Engineering, Electrical Engineering, Engineering Physics, Environmental Engineering, Industrial Engineering, Information Engineering, Mechanical Engineering, Material Science and Engineering, Manufacturing Processes, Microelectronics, Mining Engineering, Petroleum Engineering, and other application of physical, biological, chemical and mathematical sciences in engineering. Authors are invited to submit articles that have not been published previously and are not under consideration elsewhere.

Starting from Vol. 35, No. 1, 2003, full articles published are available online at <http://journal.itb.ac.id>, and indexed by Scopus (2018:Q2), Index Copernicus, Google Scholar, DOAJ, The Elektronische Zeitschriftenbibliothek EZB by University Library of Regensburg, EBSCO Open Science Directory, Ei Compendex, Chemical Abstract Service (CAS), Zurich Open Repository and Archive Journal Database, Emerging Sources Citation Index and Indonesian Publication Index.

The journal has been also accredited for five years based on The Decree of Directorate General of Research and Development Strengthening, Ministry of Research, Technology and Higher Education of the Republic of Indonesia No. 30/E/KPT/2018, for Volume 50 No.1, 2018 until Volume 54 No. 6, 2022.

ISSN: 2337-5779  
E-ISSN: 2338-5502

Published by ITB Journal Publisher, the Institute for Research and Community Services, Institut Teknologi Bandung, in collaboration with Indonesian Engineering Association (Persatuan Insinyur Indonesia-PTI).

Publication History

Formerly known as:

- ITB Journal of Engineering Science (2007 - 2012)
- Proceedings ITB on Engineering Science (2003 - 2007)
- Proceedings ITB (1961 - 2002)


Back issues can be read online at: <http://journal.itb.ac.id>

Scimago Journal Ranking



00105760

View My Stats



This work is licensed under a Creative Commons Attribution-NoDerivatives 4.0 International License.

User

Username  
Password  
 Remember me  
Login

Notifications

View  
Subscribe





## Journal Content

Search

Browse

[By Issue](#)

[By Author](#)

[By Title](#)

[Other Journals](#)

## Information

[For Readers](#)

[For Authors](#)

[For Librarians](#)

[Home](#) > [About the Journal](#) > [Editorial Team](#)

## Editorial Team

### Editor-in-Chief

[Prof. Dr. Tjandra Setiadi](#), Faculty of Industrial Technology, Institut Teknologi Bandung, Indonesia

### Editor

[Dr. T. Taufik](#), Electrical Engineering Department, California Polytechnic State University,, United States

[Dr. Tobias Nilsson](#), Geodesy and Remote Sensing, Helmholtz Centre Potsdam, GFZ German Research Centre for Geosciences,, Germany

[Dr. Judha Puradilaksono](#), Department of Mechanical Engineering, Faculty of Engineering, Universiti Teknologi Brunei, Brunei Darussalam

[Prof. Jeong Guon In](#), Center for Noise and Vibration Control, Department of Mechanical Engineering, Korea Advanced Institute of Science and Technology, Korea, Republic of

[Prof. Chien Ming Wang](#), School of Civil Engineering, University of Queensland, Australia

[Prof. M. Akbar Rhamdhani](#), Faculty of Science, Engineering & Technology School of Engineering, Swinburne University of Technology, Australia

[Dr. Tech. Ir Dudy Darmawan Wijaya](#), Geodesy Research Group ITB, Indonesia

[Prof. Dr. Satriá Bijaksana](#), Global Geophysics Research Group, Faculty Mining and Petroleum Engineering, Bandung Institute of Technology, Indonesia

[Prof. Dr. Ir. Tatacipta Dirgantara](#), Faculty of Mechanical and Aerospace Engineering, Institut Teknologi Bandung, Indonesia

[Prof. Dr. Ir. Marinus Cornelis Maria \(Mark\) van Loosdrecht](#), Department of Biochemical Engineering, Delft University of Technology, Netherlands

[Prof. Dr. Ai Qin Wang](#), Center for Eco-material and Green Chemistry, Lanzhou Institute of Chemical Physics, Chinese Academy of Sciences, P.R. China

[Prof. Dr. Ir. Masnur Isyamb](#), Faculty of Civil and Environmental Engineering, Institut Teknologi Bandung, Indonesia

[Dr. Ir. Indan Rachmatian Siti Salemi](#), Faculty of Civil and Environmental Engineering, Institut Teknologi Bandung, Indonesia

[Prof. Dr. Daddy Kurniadi](#), Faculty of Industrial Technology, Institut Teknologi Bandung, Indonesia, Indonesia

[Prof. Dr. Ir. Suwarno Suwarno](#), School of Electrical Engineering and Informatics, Institut Teknologi Bandung, Indonesia

[Dr. Shuren Wang](#), School of Civil Engineering and Mechanics, Yanshan University, Qinhuangdao, China

### Managing & Layout Editor

[Dini Sofiani Permatasari](#), The Institute for Research and Community Services, Institut Teknologi Bandung, Indonesia

### Copyeditor

[Syorand Zijlstra](#)

### Publication Assistant

[Yulian Qotiman](#), The Institute for Research and Community Services, Institut Teknologi Bandung, Indonesia

ISSN: 2338-5502



### Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)

Herio Tanujaya<sup>1</sup>, Harufumi Shiotsuka<sup>2</sup>, Dai Kitagawa<sup>3</sup>,  
Adianto<sup>4</sup>, Susukidinata<sup>4</sup> & Setyaji Kawan<sup>4</sup>

<sup>1</sup>Mechanical Engineering Department, Faculty of Engineering,  
Tartemmagata University, Jalan Lelisa, S. Paronon No. 7, Jakarta 11440, Indonesia  
<sup>2</sup>Faculty of Medicine, Tarumanagara University,  
Jalan Leleica, S. Paronon No. 1, Jakarta 11440, Indonesia  
<sup>3</sup>Mechanical Science and Biomechanical Department, Graduate School of Engineering  
Science, Osaka University, Machikaneyama-cho 1-3, Toyonaka, Osaka 560-0571, Japan  
<sup>4</sup>Email: heri.tan@tntu.ac.id

**Abstract.** In this research, we have developed, fabricated, tested, and analyzed an artificial basilar membrane prototype (ABMP), which works using sinusoidal waves of various frequencies. The design of the prototype was a trapezoidal shape with a length of 20 mm and a width of 2 to 4 mm. The research was carried out experimentally and analytically. Experimentally, the ABMP's vibration was measured using a laser Doppler vibrometer (LDV) and a function generator to generate various frequencies. The analytical approach is discussed based on the Weizel-Kramer-Billion method (WKB). The results show that resonance frequencies can be reached within the range of human hearing, between 20 Hz to 20 kHz.

**Keywords:** ABMP, cochlea, frequency, frequency selectivity, PVDF, resonance, vibrating amplitude, WKB

#### 1 Introduction

The ears are a component of the human auditory system. Hearing impairment can be brought about by many causes. Sensorimotor hearing loss is one of them. Usually, in medicine a cochlear implant (CI) or hearing aid (HA) is used to assist the patient who suffers from deafness. In this paper, we discuss an artificial basilar membrane prototype (ABMP) using a piezoelectric membrane made of polyvinylidene fluoride (PVDF). The function of the cochlea in the human auditory system is not only to convert acoustic sound to electrical signals but also frequency selectivity. The biological basilar membrane, which is a resonator in the cochlea, plays a prominent role in frequency selectivity. In this experiment, we report the development of a fully self-contained artificial cochlea, the aforementioned ABMP. The acoustic sensor which is reported in this paper realizes frequency selectivity in the air. Normal humans can hear sounds in the frequency range of 20 Hz to 20 kHz.

Received on August 8<sup>th</sup>, 2011; 1<sup>st</sup> Revision on November 1<sup>st</sup>, 2011; 2<sup>nd</sup> Revision May 2<sup>nd</sup>, 2012; 3<sup>rd</sup> Revision May 24<sup>th</sup>, 2012; 4<sup>th</sup> Revision January 09<sup>th</sup>, 2013; Accepted for publication January 17<sup>th</sup>, 2013.  
Copyright © 2013, Published by ITB Research Publication, ISSN 2337-9779, DOI: 10.53114/itb.arts.v45i1.61-72

#### Match Overview

6%

- 6
- 1
- 2
- 3
- 4
- 5
- 6

1	journals.itb.ac.id	4%
2	Submitted to Universiti	2%



## ORIGINALITY REPORT

---

**6%**

SIMILARITY INDEX

**4%**

INTERNET SOURCES

**0%**

PUBLICATIONS

**2%**

STUDENT PAPERS

---

## PRIMARY SOURCES

---

**1**

**journals.itb.ac.id**

Internet Source

**4%**

---

**2**

**Submitted to Universiti Teknikal Malaysia  
Melaka**

Student Paper

**2%**

---

Exclude quotes  On

Exclude bibliography  On

Exclude matches  < 2%

# ITB

*by* Fakultas Teknik

---

**Submission date:** 06-Dec-2019 11:59AM (UTC+0700)

**Submission ID:** 1228349090

**File name:** Experimental\_and\_analytical\_study.pdf (1.41M)

**Word count:** 3888

**Character count:** 19130



## 1 Experimental and Analytical Study Approach of Artificial Basilar Membrane Prototype (ABMP)

Harto Tanujaya<sup>1</sup>, Hirofumi Shintaku<sup>3</sup>, Dai Kitagawa<sup>3</sup>,  
Adianto<sup>1</sup>, Susilodinata<sup>2</sup> & Satoyuki Kawano<sup>3</sup>

<sup>1</sup>Mechanical Engineering Department, Faculty of Engineering,  
Tarumanagara University, Jalan Letjen. S. Parman No. 1, Jakarta 11440, Indonesia

<sup>2</sup>Faculty of Medicine, Tarumanagara University,  
1 Jalan Letjen. S. Parman No. 1, Jakarta 11440, Indonesia

<sup>3</sup>Mechanical Science and Bioengineering Department, Graduate School of Engineering  
Science, Osaka University, Machikaneyama-cho 1-3, Toyonaka, Osaka 560-8531, Japan  
Email: hart\_tan18@yahoo.com

**1 Abstract.** In this research, we have developed, fabricated, tested, and analyzed an artificial basilar membrane prototype (ABMP), which works using sinusoidal waves of various frequencies. The design of the prototype has a trapezoidal shape with a length of 30 mm and a width of 2 to 4 mm. The research was carried out experimentally and analytically. Experimentally, the ABMP's vibration was measured using a laser Doppler vibrometer (LDV) and a function generator to generate various frequencies. The analytical approach is discussed based on the Wentzel Kramer Brillouin method (WKB). The results show that resonance frequencies can be reached within the range of human hearing, between 20 Hz to 20 kHz.

**Keywords:** ABMP; cochlea; frequency; frequency selectivity; PVDF; resonance; vibrating amplitude; WKB.


### 1 Introduction

The ears are a component of the human auditory system. Hearing impairment can be brought about by many causes. Sensorineural hearing loss is one of them. Usually, in medicine a cochlear implant (CI) or hearing aid (HA) is used to assist the patient who suffers from deafness. In this paper, we discuss an artificial basilar membrane prototype (ABMP) using a piezoelectric membrane made of polyvinylidene fluoride (PVDF). The function of the cochlea in the human auditory system is not only to convert acoustic sound to electrical signals but also frequency selectivity. The biological basilar membrane, which is a resonator in the cochlea, plays a prominent role in frequency selectivity. In this experiment, we report the development of a fully self-contained artificial cochlea, the aforementioned ABMP. The acoustic sensor which is reported in this paper realizes frequency selectivity in the air. Normal humans can hear sounds in the frequency range of 20 Hz to 20 kHz.



## 2 Mechanical Model

For modeling purposes and in order to simplify the model of the biological basilar membrane, the ABMP can be unwrapped. There are two channels above and below the ABMP, as shown in Figure 1(a). The channels are considered as the scala vestibule and the scala tympani in the biological cochlea. Figure 1(a) shows the complete geometry of the ABMP model, from the top and through its cross section. This condition is made for the analytical calculation. Figure 1(b) shows a 3D view of the ABMP that was used in our experiment.



**Figure 1** Geometric model of artificial basilar membrane prototype (ABMP), (a) top and cross section view, (b) 3D view.

The ABMP is made of PVDF membrane and is bonded on a stainless steel plate, which has a trapezoidal shaped perforation with a width ranging from 2 mm to 4 mm and a length of 30 mm. The thickness of the ABMP is 40  $\mu\text{m}$ . There are 24 detecting electrodes on the ABMP, which are fabricated using photolithography and etching. The whole array of detecting electrodes is made of aluminium thin film with a thickness of 500 nm and they are located on the central line of the trapezoidal shape. The Young modulus and density of the membrane are 4 GPa and 1790 kg/m<sup>3</sup> respectively. The membrane and the perforated trapezoid are mounted on a substrate with a channel. The channel's dimensions are 47  $\times$  17 mm (rectangular) and a depth of 4 mm, as shown in Figure 1(b).

## 3 Analytical Approach

Many researchers have analyzed the theory of the cochlea. Von Békésy [1] analyzed a mechanical cochlea to study the traveling wave. Steele & Taber [2] analyzed and constructed a mechanical cochlear model using a three-dimensional model. Zhou, *et al.* [3] used an isotropic polymer membrane placed

on a plate to build the first life-sized mechanical cochlea. Kirchhoff's hypotheses are fundamental assumptions for the development of linear, elastic, small-deflection theories concerning the bending of thin plates [4]. Thin plate bending theory usually refers to Kirchhoff's plate theory. The method of Wentzel Kramer Brillouin (WKB) was also used to solve the equations of the mechanism of the membrane problem, whereby the parameters are very slowly compared to the wavelength of the oscillations. The solution of the vibration of the membrane response is assumed with a slowly varying amplitude. Interaction between the fluid and the membrane is governed by linearized Euler equations.

In order to obtain the oscillatory solution at the steady state, some assumptions are made. The following assumptions are used to derive the basic equations of the ABMP. Vibration as a thin plate can be satisfied under the following conditions: deflections of the ABMP are very small compared to its thickness and the tensile term is negligible in the response of the structure. According to these assumptions, the ABMP is modeled as a thin bending plate, because the thickness of the membrane is extremely small compared to its width and length, and the plane stress is assumed to be  $\sigma_z = \tau_{xz} = \tau_{yz} = 0$ . Based on the shearing force and bending moment of the ABMP model, and the equation of motion, the governing equation of the bending vibration of the plate with isotropic mechanical properties can be described as,

$$D \left( \frac{\partial^4 w}{\partial x^4} + 2 \frac{\partial^2 w}{\partial x^2 \partial y^2} + \frac{\partial^4 w}{\partial y^4} \right) + \rho h \frac{\partial^2 w}{\partial t^2} = 0 \quad (1)$$

where,

$$D = \frac{Eh^3}{12(1-\nu^2)} \quad (2)$$

E : Young Modulus

h : thickness of the membrane

$\nu$  : Poisson's ratio.

Problems for fluid structure interaction are relatively complex, so to obtain a mathematical solution, the following assumptions are made, based on the experimental observations. The ABMP can be unrolled in order to simplify the model and the vibration of the ABMP is modeled as the vibration of a thin plate with small bending amplitude; the fluid flow is assumed to be incompressible, and the gravity effects of the surrounding fluid were ignored. Furthermore, the following assumptions are made in order to obtain the oscillatory solution at the periodic steady state. A single mode  $\eta(x, y)$  is used for the shape function of the

ABMP's bending in  $y$  direction.  $\eta(x, y)$  is determined based on the analytical solution of a vibrating beam with the first mode, the length of  $b(x)$ , and the fixed boundary conditions at  $y = \pm \frac{b(x)}{2}$  as,

$$\eta(x, y) = \begin{cases} c_1 \cos\left(\frac{\beta}{b(x)} y\right) + c_2 \cosh\left(\frac{\beta}{b(x)} y\right) & \text{at } -\frac{b(x)}{2} \leq y \leq \frac{b(x)}{2} \\ 0 & \text{at } -\frac{L_1}{2} \leq y \leq -\frac{b(x)}{2} \text{ and } \frac{b(x)}{2} \leq y \leq \frac{L_1}{2} \end{cases}, \quad (3)$$

where  $c_1$ ,  $c_2$ , and  $\beta$  are constants, 0.8827, 0.1173 and 4.730, respectively, which are obtained and compared with the experimental results. These constants are determined so as to make  $\eta(x, y)$  satisfy the fixed boundary conditions at  $y = \pm b(x)/2$ .

The wave is considered as a slowly varying wave in  $x$  direction [5]. That is, wave number  $k(x)$  is slowly varying along  $x$  as  $b(x)$ , where  $db(x)/dx \approx 0$  and  $dk(x)/dx \approx 0$  are reasonable in the scale of one wavelength. In this case, the waves can be treated as pseudo plane waves and can be described by the WKB asymptotic solution [4]. Based on the assumptions described above, the displacement  $w(x, y, t)$  of the ABMP can be written as

$$w = W(x)\eta(x, y)e^{i\int_0^x k(\xi)d\xi} e^{-i\omega t}, \quad (4)$$

where  $i$  and  $W(x)$  are the imaginary number and the envelope function, respectively.  $W(x)$  is also treated as a slowly varying function, that is  $dW(x)/dx \approx 0$ , since the effect of  $dW(x)/dx$  on the dispersion relationship is trivial for linear problems. The basic equation of the fluid flow is the Laplace equation of velocity potential  $\phi_i$ , written as

$$\frac{\partial^2 \phi_i}{\partial x^2} + \frac{\partial^2 \phi_i}{\partial y^2} + \frac{\partial^2 \phi_i}{\partial z^2} = 0. \quad (5)$$

The boundary conditions of the normal velocities at the wall of the fluid channel are written as

$$u_z = \frac{\partial \phi_i}{\partial z} = 0 \text{ at } z = -L_2, \quad (6)$$



$$u_y = \frac{\partial \phi_l}{\partial y} = 0 \text{ at } y = \pm \frac{L_1}{2}, \quad (7)$$

The subscript  $f$  is  $u$  or  $l$ , where  $u$  and  $l$  indicate the fluid at the upper and lower sides of the ABMP, respectively.  $L_1$  and  $L_2$  are the width and the depth of the fluid channel, respectively. The kinematic boundary condition at  $z=0$  is written as

$$\frac{\partial w}{\partial t} = \frac{\partial \phi_f}{\partial z} \text{ at } z = 0 \quad (8)$$

Therefore, the velocity potential can be written as

$$\phi_f = \sum_{j=0}^{\infty} A_j \cosh[\zeta_j(z + L_2)] \cos\left(\frac{j\pi}{L_1} y\right) e^{\int_0^z k(\xi) d\xi} e^{-i\omega t}, \quad (9)$$

where  $A_j$  and  $\zeta_j$  are the Fourier coefficient for the  $j$ th mode and  $[k^2(x) + (j\pi/L_1)^2]^{1/2}$ , respectively. From Eqs. (4), (8), and (9), the following equation is obtained:

$$i\omega W(x)\eta(x, y) = -\sum_{j=0}^{\infty} A_j \zeta_j \sinh(\zeta_j L_2) \cos\left(\frac{j\pi}{L_1} y\right). \quad (10)$$

Using the orthogonality of the cosine function,  $A_j$  is calculated as

$$A_j = -\frac{i\omega W(x) \int_{-b(x)/2}^{b(x)/2} \eta(x, y) \cos(j\pi y / L_1) dy}{\zeta_j \sinh(\zeta_j L_2) \int_{-L_1/2}^{L_1/2} \cos^2(j\pi y / L_1) dy}. \quad (11)$$


Using the boundary of integration from  $-b(x)/2$  to  $+b(x)/2$  with respect to  $y$ , the following *eikonal equation* is obtained:

$$\begin{aligned} & f(x, \omega) \\ &= D \left[ k^4(x) \int_{-b(x)/2}^{b(x)/2} \eta^2(x, y) dy \right. \\ & \quad - 2k^2(x) \int_{-b(x)/2}^{b(x)/2} \eta(x, y) \partial^2 \eta(x, y) / \partial y^2 dy \\ & \quad \left. + [\beta / b(x)] \int_{-b(x)/2}^{b(x)/2} \eta^2(x, y) dy \right] \\ & \quad - \omega^2 \left[ \rho_m h \int_{-b(x)/2}^{b(x)/2} \eta^2(x, y) dy + \rho_1 \sum_{j=0}^{\infty} \frac{\left[ \int_{-b(x)/2}^{b(x)/2} \eta(x, y) \cos(j\pi y / L_1) dy \right]^2}{\zeta_j \tanh(\zeta_j L_2) \int_{-L_1/2}^{L_1/2} \cos^2(j\pi y / L_1) dy} \right] = 0. \end{aligned} \quad (12)$$


The eikonal equation describes the dispersion relationship between  $k(x)$  and  $\omega$  at various  $x$ . The effect of the surrounding fluid is found in the last term of this equation. Since this term contributes to increasing the effective mass of the vibration, the resonant frequency may be decreased by the surrounding fluid. From the average variation principle, it is known that the *eikonal equation* has a relationship with  $W(x)$ :

$$W(x) = \frac{c}{\left(\frac{\partial f}{\partial k}\right)^{1/2}}, \quad (13)$$

where  $c$  is a constant. Eq. (13) is the *transport equation* which describes the qualitative distribution of  $W(x)$ . Figure 2 shows the graph of envelope curve  $W(x)$  in the air environment defined by Eq. (13). In that figure, the peaks at certain frequencies indicate local resonance at a particular place of the ABMP. The peaks of  $W(x)$  at each frequency show changing to the smaller  $x$  with increasing the frequency of the acoustic wave.



**Figure 2** Envelope curve  $W(x)$  at 6, 9, and 12 kHz, which is determined on the basis of the analytical solution.



**Figure 3** Description of the envelope curve  $W(x)$  and a single mode of the vibrating amplitude in  $y$  direction  $\eta(x, y)$ .

Figure 3 shows the description of the envelope curve  $W(x)$  and a single mode of the vibrating amplitude in y direction  $\eta(x, y)$  at a certain frequency. The maximum vibrating amplitude at a certain frequency  $L2$  occurs at the same place of the local resonance frequency.

#### 4 Experimental Setup

Experiments to realize frequency selectivity have been reported by some researchers. Wittbrodt, *et al.* [6] developed a membrane made of polyimide using aluminum beams. Their paper states that this membrane had traveling waves similar to those of the biological basilar membrane. Chen, *et al.* [7] made a membrane by depositing discrete Cu beams on a piezo membrane. They also simulated a membrane using a computational model of the mechanical device. Tanaka, *et al.* [8] developed an acoustic sensor using “fishbone” method as a cantilever arrays which realized the frequency selectivity. White, *et al.* [9] developed a membrane made of polyimide with  $\text{Si}_3\text{N}_4$  beams.


There are 24 electrodes on the membrane of the ABMP. These electrodes are used to measure the membrane voltage. We measured the membrane vibration along the whole membrane at the middle of each electrode. Figure 4 shows a diagram of the experimental setup: (a) 3D view and (b) experimental equipment. A sinusoidal acoustic wave was applied to the ABMP via a speaker (FOSTEX, Japan). Previously, the speaker was calibrated to realize a constant sound pressure at various frequencies. The function generator (NF, Japan) connected to the speaker was used to control the frequency of the sound wave of the speaker from 1 to 20 kHz. The speaker was connected in parallel with a signal/function generator and a computer, as shown in Figure 4. We used VBA programming and a UDF, which was inserted in the Excell program to drive the various applied frequencies and sound amplitude to the speaker automatically. We used those frequencies that are within the range of the human audible frequency. An oscilloscope was connected to the speaker in order to analyze the output data sampling at the same time with the input of the various applied frequencies and vibration amplitude of the membrane. The VBA programming and UDF were used to arrange and save the input/output data from a signal/function generator and an oscilloscope. The distance between speaker and membrane was 120 mm. The voltage of the speaker was also adjusted, using a function generator to get a constant sound pressure of 75 dB.

A photograph of the membrane developed in this research is shown in Figure 4(b). The membrane is bonded on the plate with the perforated trapezoid, such that the membrane can easily vibrate by applying an acoustic wave. Although the experiment was done in the atmosphere, both sides of the biological basilar membrane were in lymph fluid. However, the main purpose of this paper is to



understand the basic mechanism of frequency selectivity. In this research, we have used a prototype membrane made of PVDF. The affect and position of the excitation of the membrane are needed to develop a membrane made of P(VDF-TrFE) in the future.

The ABMP was placed on a motorized stage, which could be moved in  $x$  and  $y$  directions. The motorized stage was used to examine the spatial distribution of the vibrating amplitude on the surface of ABMP automatically. The vibrating amplitude of the ABMP in  $z$  direction was measured by a laser Doppler vibrometer (LDV) (Graphtec, Japan). Using the LDV, the mechanical vibration of the ABMP was converted to velocity data. The velocity data sampling was later analyzed by a Fast Fourier Transform (FFT) using the oscilloscope to obtain a vibration amplitude at the human audible frequency of acoustic waves.




**Figure 4** Experimental setup: (a) 3D view and (b) experimental equipment.

The ABMP is relatively large to be directly implanted into the biological cochlea. This problem can be solved by using a micro fabrication method. Optimization and minimization of the membrane remain to be done in a future work.

## 5 Results and Discussions

The first experiment was done in the atmosphere in order to learn the characteristics of the frequency selectivity mechanism of the prototype membrane. Figure 5 shows a contour map of the vibrating amplitude of the ABMP membrane at  $f =$  (a) 6 kHz, (b) 9 kHz, and (c) 12.8 kHz, respectively. These contour maps show the different positions of the maximum amplitude at each frequency. Position  $x$  of the maximum amplitudes decreased as the frequency increased.



**Figure 5** Contour map of the vibrating amplitude of the ABMP membrane at  $f =$  (a) 6 kHz, (b) 9 kHz, and (c) 12.8 kHz.

The local maximum amplitudes are considered to be the result of the standing waves in  $x$  direction. The maximum vibrating amplitude occurred at the distances  $x =$  (a) 27.5 mm, (b) 19 mm, and (c) 8.5 mm from the base, with an amplitude of 36.7 nm, 35 nm, and 54.6 nm, respectively.


Several small peaks were observed at a frequency area lower than the resonance frequency, which was caused by the standing wave. The standing wave was not observed at frequencies higher than the resonating place, which was caused by the fact that the wavelength is relatively long at higher frequencies. Referring to the theory, it is confirmed that the wavelength  $k(x)$  is relatively small. In the biological cochlea, the acoustic wave travels from the basal (wider area) to the apex (sharper area). This condition shows that the characteristics of the wave and the oscillation of the prototype membrane are the same as those of the biological membrane.

Figure 6 shows the frequency-dependence of the vibrating amplitude of the ABMP at the distances  $x =$  (a) 27.5 mm, (b) 19 mm, and (c) 8.5 mm. The maximum vibrating amplitudes show clear peaks at different frequencies. The frequency at the maximum vibrating amplitude is considered to be the resonant frequency at that local area of the membrane. The maximum frequency increases as the  $x$  position decreases. This corresponds to the results in Figure 5. That is, the wavelength of the acoustic wave was affected by the width of the membrane. When the ABMP is relatively wide in  $x$  direction compared to the width in  $y$  direction, the vibration is mainly effected by the boundary conditions at  $y = \pm b(x)/2$ . Oscillation of the membrane is dominated by the ABMP's local


structural strain in  $y$  direction. Furthermore, the vibration in  $y$  direction is larger than that in  $x$  direction.

Figure 7 shows the relationship between  $x$  the length of the membrane and the local resonance frequency,  $f_l$ , at 75 dB. The local resonance frequency,  $f_l$ , has a value similar to that of frequency  $f$ . Experimentally, the frequency selectivity of the membrane output is induced by the local resonance of the displacement of the membrane. The local resonance frequency,  $f_l$ , decreases from 14 to 3 kHz with the increase of position  $x$  of the membrane length. The local resonance frequencies of the membrane at the positions of  $x$  were outside of the range of the resonance frequency linetheory. Almost all experimental results were slightly lower than the predictions. Perhaps this is due, firstly, to the fabrication process (effects from photolithography and etching), after which the density of the membrane has changed, and secondly, to an underestimation of Young modulus  $E$ . In the literature,  $E$  is widely distributed as 3.0~11.0 GPa.

The resonance frequencies of the theoretical analysis predict that the frequency range can be controlled by changing the thickness, density, Young modulus and length of the membrane. Figure 7 also shows frequency  $f$  compared with the theoretical resonance frequency,  $f_l$ , based on the plate bending theory that is solved with the WKB asymptotic method. Agreement between the experimental results and the theoretical analysis was confirmed.



**Figure 6** Frequency dependence of the vibrating amplitude of the ABMP at the distances  $x =$  (a) 27.5 mm, (b) 19 mm, and (c) 8.5 mm.



**Figure 7** Relationship between the length of the membrane ( $x$ ) and the local resonance frequency ( $f$ ) at 5 dB.

## 6 Conclusions

We have presented an analytical and experimental investigation of our artificial basilar membrane prototype (ABMP). Frequency selectivity of the membrane is confirmed at the range of 3 ~ 14 kHz. The resonant frequency increases as the width of the membrane decreases. Using theoretical calculation, we can predict the thickness of a suitable membrane to be fabricated. Agreement between the experimental data and the theoretical analysis was confirmed. Some of the experimental results were not the same as, or out of the range of the theoretically predicted values. This can be explained by the fabrication process of the membrane and underestimation of the Young Modulus  $E$ .

Future work is needed to make the artificial cochlea better as an acoustic sensor. The phase of the oscillations was neglected in this research, since the goal in this experiment was to confirm the frequency selectivity of the prototype membrane. The phase can refer to a specific references of wave and it will be useful to use P(VDF-TrFE) as the material in the next work. The voltage of the artificial auditory membrane is our end goal. The effect of viscosity may also contribute to improve the frequency selectivity of the membrane. Material and fabrication process influence the characteristics of the membrane. We need to minimize the membrane, because the present membrane is relatively large for implantation into a cochlea. A thinner membrane has a vibration amplitude larger than a thicker one, it is relation with the thinner membrane can generate the bigger voltage.

### Acknowledgements

The author would like to thank the reviewers for their relevant comments, and special thanks are due to Dr. Yoichi Kagaya and all members of the Kawano Laboratory at the Mechanical Science and Bioengineering Department, Osaka University, Japan.

### References

- [1] Békésy, von G., *Current Status of Theories of Hearing*, Science, **123**(3201), pp. 779-783, 1956.
- [1] Békésy, von G., *Experiments in Hearing*, McGraw Hill, New York, 1960.
- [2] Steele, Charles R. & Taber, Larry A., *Comparison of WKB Calculations and Experimental Results for Three-Dimensional Cochlear Models*, J. Acoustical Society of America, **65**(4), April 1979.
- [3] Zhou, G., Bintz, L., Anderson, D.Z. & Bright, K.E., *A Life-sized Physical Model of the Human Cochlea with Optical Holographic Readout*, J. Acoust. Soc. Am., **93**(3), pp.1516-1523, 1994.
- [4] Ventzel, Eduard & Krauthammer, Theodor, *Thin Plates and Shells: Theory, Analysis, and Applications*, Marcel Dekker, Inc., New York, USA, 2001.
- [5] Whitham, G.B., *Linear and Nonlinear Waves*, Wiley-Interscience, 1974.
- [6] Wittbrodt, M., Puria, S. & Steele, C.R., *Developing a Physical Model of the Human Cochlea Using Micro-fabrication Methods*, Audiology and Neurotology, **11**(2), pp.104-112, 2006.
- [7] Chen, F., Cohen, H., Bifano, T.G., Castle, J., Fortin, J., Kapusta, C., Mountain, D.C., Zosuls, A. & Hubbar, A.E., *A Hydromechanical Biomimetic Cochlea: Experiments and Models*, J. Acoust. Soc. Am., **119**(1), pp. 394-405, 2006.
- [8] Tanaka, K., Abe, M. & Ando, S., *A Novel Mechanical Cochlea "fishbone" with Dual Sensor/Actuator Characteristics*, IEEE/ASME Transactions on Mechatronics, **3**(2), pp.98-105, 1998.
- [9] White, R.D. & Grosh, K., *Microengineered Hydromechanical Cochlear Model*, Proceedings of the National Academy of Sciences of the United States of America, **102**(5), pp.1296-1301, 2005.